



CERTIFICATE OF COMPLIANCE

Certificate No.:

C183810.01

Date of Issue:

16 July 2018

Applicant Name:

Sion Technologies LTD.

Applicant Address:

Shprintzak 12A, K.Ata, 2806516, Israel

Product:

Hydrogen carrier gas solution for GC/GCMC systems

Model / Type Ref.

UCS-1000

Rating:

**100/240 VAC, 0.5A, 50/60 Hz,
Gas supplies: H₂, Air, He/N₂: 40-90 psi.**

Tested to:

EN 61010-1 (Third Edition): 2010;

**Test Results are detailed in
the Test Report No.**

S183810.01

Name: **Yigal Cohen**

Signature: _____

Position: **ITE Team leader**

Department: **Product Safety**

This is to certify that the product specified herein has been tested & the test results were found to be compliant with the requirements noted above.